

CERTIFICATE OF CALIBRATION

Date Of Issue
6th January 2021

Certificate Number
060121/N20



No. 0601

Certificate Issued By:



Electronic Temperature Instruments Ltd
Easting Close, Worthing
West Sussex BN14 8HQ, UK
tel: 01903 202151
email: sales@etiltd.co.uk
www.etiltd.com

Page 1 of 2 pages	
Approved Signatory	Signature
S. Saunders	

Customer Name: SIGNATROL LIMITED
Address: UNIT E2
GREEN LANE BUSINESS PARK
TEWKESBURY
GLOUCESTERSHIRE
GL20 8SJ

Order Number: 46165

Ref Number: 839/83785

Date Received: 18th December 2020

Date Calibrated: 6th January 2021

Ambient Temperature: 23 °C ± 3 °C

Ambient Humidity: < 60 %rh

Temperature Scale: International Temperature Scale of 1990

Inst Description: ETI REFERENCE THERMOMETER

Sensor Type: INTEGRAL 130mm PT100 LIQUID PROBE

Inst Serial Number: D12360142

Probe Serial Number: N/A

Calibrated Range/Scale: - 50 °C to + 134 °C / 0.01 °C RESOLUTION

Procedure: The Instrument was stabilised at ambient temperature, then calibrated by comparison with two traceable references in stirred liquid baths. Immersion depth 150mm

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

The results indicate performance of instrument at time of measurement, with no warranty as to specification, repeatability or long term stability.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No 0601

Certificate Number

060121/N20

Page 2 of 2 Pages

Instrument Serial Number: D12360142

Sensor Serial Number: N/A

Results:

Test Temperature °C	Instrument Reading °C
---------------------	-----------------------

0.00	- 0.01
- 49.99	- 49.97
- 18.00	- 18.02
0.00	- 0.01
4.00	4.00
25.00	25.00
100.00	100.01
134.00	134.00
0.00	0.00

Uncertainty of Measurement ± 0.04 °C + Instrument Resolution

End of Report