

CERTIFICATE OF CALIBRATION

ISSUED BY The Roxspur Measurement & Control
Calibration Laboratory



0043

Page 1 of 2

Assessed Signatory

SHAUN BOLDY



2 Downgate Drive
Sheffield
South Yorkshire
S4 8BT

t: 0114 224 9205
f: 0114 224 9224

e: service@roxspur.com
i: www.roxspur.com

Date of Issue: 05 June 2019

Certificate Number 158314

Customer: SIGNATROL LIMITED
UNIT E2
GREEN LANE BUSINESS PARK
GLOUCESTERSHIRE
GL20 8SJ

Date Received 30 May 2019
RM&C Order Ref. L508944
Customer Order No. 45789
Calibration Date 05 June 2019
Next Calibration Due 05 June 2020

Equipment Information

Description	LEYRO LDT-2000 PRECISION THERMOMETER & PT100 PROBE		
Manufacturer	LEYRO	Serial Number	1031401205 & 351839-1
Model Number	LDT-2000 & 935-14-116	Customer Inventory No.	CE1113 & CE1115
Calibrated Range	-70 °C to 300 °C	RM&C I.D. No.	RMC0044158
Scale / Resolution	0.001 °C		
Calibration Points	-70 °C, 0 °C, 30 °C, 150 °C & 300 °C		

Conditions

Lab Temperature	21.0 °C ± 2 °C	Department	TEMP - BATH
Probe Type	Pt100	Engineer	SHAUN BOLDY
Probe Length	350 mm	Last Certificate Number	135683
Probe Diameter	6 mm		
Min. Immersion Depth	200 mm		

Procedure : RM&C 023 DTI & RTD

RM&C 023: Digital Thermometer & RTD Probe – Issue 4 (Dec-2018)

The thermometer under test was allowed to equilibrate within a controlled, stable environment, the temperature of which was measured using traceable reference Platinum Resistance Thermometers. The following results indicate the measured test thermometer temperature against the measured temperature at the time of calibration. The measurement uncertainty was calculated in accordance with M3003 (Edition 3 – November 2012) and as such takes into account such factors as the calibration & drift of the reference standards, stability, repeatability and resolution of reference instruments and that of the unit under test.

The results are valid at the time of calibration only. The temperature scale used was ITS-90. All measurements are traceable to National Standards. Calibration has been carried out using Laboratory procedures (LAB-PROC-023) in accordance with BS EN ISO 17025. The results are valid at the time of calibration only and are "As Found" (i.e. No Adjustments Made).

Notes :

No measured errors, in the parameters checked, exceeded the customer specified tolerance of ± 0.04 °C, not taking into account the uncertainty of measurement.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0043

Certificate Number

158314

Page 2 of 2

Calibration Results

Serial No: 1031401205 & 351839-1

Reference Temperature °C	Thermometer Reading °C	Measured Error °C	Measurement Uncertainty ± °C
-0.003	0.001	0.004	0.06
-70.039	-70.056	-0.017	0.06
-0.001	-0.004	-0.003	0.06
30.052	30.064	0.012	0.06
150.085	150.104	0.019	0.06
299.855	299.871	0.016	0.06
0.001	0.007	0.006	0.06

- The certificate of calibration only applies to the instrument(s) listed on page one of the certificate -
- End of Certificate -

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor of $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. The certificate may not be reproduced other than in full, except with the prior written approval of the issuing Laboratory.