

# CERTIFICATE OF CALIBRATION


Date Of Issue  
15th November 2017

Certificate Number  
151117/N28



0601

Certificate Issued By:



Electronic Temperature Instruments Ltd  
Easting Close, Worthing  
West Sussex BN14 8HQ, UK  
tel: 01903 202151  
email: sales@etiltd.co.uk  
www.etiltd.com

Page 1 of 1 pages

Approved Signatory

Signature

J.S. Carswell

Customer Name: SIGNATROL LIMITED  
Address: UNIT E2  
GREEN LANE BUSINESS PARK  
TEWKESBURY  
GLOUCESTERSHIRE  
GL20 8SJ

Order Number: 45464

Ref Number: 617/61568

Date Received: 13th November 2017

Date Calibrated: 15th November 2017

Ambient Temperature: 22 °C ± 2 °C      Ambient Humidity: < 60 %rh

Temperature Scale: International Temperature Scale of 1990

Inst Description: ETI REFERENCE THERMOMETER

Sensor Type: INTEGRAL 130mm PT100 LIQUID PROBE

Inst Serial Number: D12360142

Sensor Serial Number: N/A

Calibrated Range/Scale: - 50 °C to + 134 °C / 0.01 °C RESOLUTION

Procedure: The Instrument was stabilised at ambient temperature, then calibrated by comparison with two traceable references in stirred liquid baths. Immersion depth 110mm

Results: Test Temperature °C      Instrument Reading °C

- 50.00	- 50.02
- 18.00	- 18.01
4.00	3.99
25.00	25.00
100.00	100.01
133.99	134.01

Uncertainty of Measurement ± 0.04 °C + Instrument Resolution.

End of Report

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%.*

*The uncertainty evaluation has been carried out in accordance with UKAS requirements.*

Results indicate performance of instrument at time of measurement, with no warranty as to specification, repeatability or long term stability.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/ or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes, including the National Institute of Standards and Technology (NIST).

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.