

CERTIFICATE OF CALIBRATION

Date Of Issue
28th September 2010

Certificate Number
280910/N15



0601

Certificate Issued By:



Electronic Temperature Instruments Ltd
Easting Close, Worthing
West Sussex BN14 8HQ England
Telephone +44 01903 202151
Facsimile +44 01903 202445
email: sales@etiltd.co.uk
Website: www.etiltd.co.uk

Page 1 of 2 pages

Approved Signatory

Signature

J. Carswell

J. McKernan

D. Carter

Customer Name: SIGNATROL LIMITED

Address: 1st & 2nd FLOOR OFFICES
105 CHURCH STREET
TEWKESBURY
GLOUCESTERSHIRE
GL20 5AB

Order Number: 43564

Ref Number: 257/25536

Date Received: 21st September 2010

Date Calibrated: 23/24th September 2010

Ambient Temperature: 22°C ± 2°C

Ambient Humidity: <60% rh

Temperature Scale: International Temperature Scale of 1990

Instrument Type : EUROTRON MICROCAL 1 SIMULATOR

Instrument Serial Number: 0049272

Procedure: The instrument was stabilised at ambient temperature, then calibrated by measuring its output on traceable reference equipment.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Results indicate performance of instrument at time of measurement, with no warranty as to specification, repeatability or long term stability.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurements to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No 0601

Certificate Number
280910/N15


Page 2 of 2 pages

Instrument Serial No 49272

Results

<u>Type "J" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
-100.0	-4.633	-4.638 8	-100.1	± 0.25 °C
700.0	39.132	39.127 1	899.9	± 0.15 °C
<u>Type "K" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	
-100.0	-3.554	-3.556 8	-100.1	± 0.25 °C
900.0	37.326	37.326 5	900.0	± 0.15 °C
<u>Type "T" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	
-100.0	-3.379	-3.371 6	-99.8	± 0.25 °C
300.0	14.862	14.860 8	300.0	± 0.15 °C
<u>Type "S" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	
800.0	7.345	7.342 6	799.8	± 1 °C
1400.0	14.373	14.371 2	1399.9	± 1 °C
<u>Type "R" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	
800.0	7.950	7.958 1	799.8	± 1 °C
1400.0	16.040	16.035 5	1399.7	± 1 °C
<u>Type "E" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	
-100.0	-5.237	-5.246 2	-100.1	± 0.25 °C
650.0	49.116	49.113 3	650.0	± 0.15 °C
<u>PT100 Output Temperature °C</u>	<u>Nominal Ohms</u>	<u>Measured Ohms</u>	<u>Equivalent Temperature °C</u>	
0.0	0.00	100.041 1	0.11	± 0.025 °C
600.0	313.71	313.661 2	599.85	± 0.18 °C

End of Report



CERTIFICATE OF CALIBRATION

ELECTRONIC TEMPERATURE INSTRUMENTS LTD

Dominion Way
Worthing
West Sussex
BN14 8NW
Tel +44 01903 202151
Fax +44 01903 202445
Email: sales@etiltd.co.uk

Customer: SIGNATROL LTD
1st & 2nd FLOOR OFFICES
105 CHURCH STREET
TEWKESBURY
GLOUCESTERSHIRE
GL20 5AB

CALIBRATION REPORT No: 280910/15
Model : EUROTRON MICROCAL 1
Standard: PREMA 6001
Date: 28th September 2010
Ambient Temperature: 22 °C ± 2 °C

Serial No: 0049272
Certificate No: CN120565
Calibrated by: J. Carswell
Ambient Humidity: < 60 %rh

Range	Test Value mV	Actual	Error	Test Limits mV
100mV	10.0	9.999	-0.001	0.006
100mV	50.0	50.008	0.008	0.006
1000mV	100.0	99.978	-0.022	0.006
1000mV	500.0	500.031	0.031	0.006
Range	Test Value mA	Actual	Error	Test Limits mA
20mA	4.0	3.9996	-0.0004	0.005
20mA	10.0	9.9985	-0.0015	0.005
20mA	20.0	19.9961	-0.0039	0.005

This is to certify that the above instrument has been calibrated against laboratory standards, which are traceable via International Agreement, to all major National Standards, including the NPL and NIST.

Certified by: 