

# CERTIFICATE OF CALIBRATION

ISSUED BY ROTRONIC INSTRUMENTS (UK) LTD



DATE OF ISSUE: 23<sup>rd</sup> June 2017

CERTIFICATE NUMBER: 30342

**rotronic**

MEASUREMENT SOLUTIONS

Calibrated by: P Image / C Aicken

Approved Signatory: M Smith 

Crompton Fields, Crompton Way,  
Crawley, West Sussex, RH10 9EE.

Telephone: 01293 571000

Fax: 01293 571008

Email: [service@rotronic.co.uk](mailto:service@rotronic.co.uk)

[www.rotronic.co.uk](http://www.rotronic.co.uk)

Page 1 of 1

Dates Measurements Performed:

21<sup>st</sup> to 23<sup>rd</sup> June 2017

Customer Details	: Signatrol Limited, Unit E2, Green Lane Business Park, Tewksbury, GL20 8SJ
Customer's Order Number	: 45383
Rotronic Ref Number	: 30342
Instrument Description	: Humidity & Temperature Probe
Manufacturer	: Rotronic AG
Model Type (s)	: HC2-S
Serial Number (s)	: 60785 619

## Humidity Procedure RUKP20

The hygrometer was calibrated using ROTRONIC non-saturated salt relative humidity (RH) standards, certified as traceable to National Standards. The probe was subjected to the relative humidity generated by the RH standard inside a calibration chamber, and the values taken from HW4. The calibration was conducted in controlled laboratory conditions 23 °C ± 2 °C.

### AS-FOUND RESULTS

Applied Relative Humidity (%rh)	Calibration Uncertainty ** (%rh)	Indicated Relative Humidity (%rh)	Instrument Error (%rh)	Indicated Temperature (°C) *	Ambient Temperature (°C) *
11.1	±0.5	10.0	-1.1	21.7	21.4
49.6	±1.1	48.7	-0.9	21.6	21.3
75.1	±1.3	74.2	-0.9	21.6	21.2

### POST-ADJUSTMENT RESULTS

Applied Relative Humidity (%rh)	Calibration Uncertainty ** (%rh)	Indicated Relative Humidity (%rh)	Instrument Error (%rh)	Indicated Temperature (°C) *	Ambient Temperature (°C) *
11.2	±0.5	11.2	0.0	22.0	21.8
49.7	±1.1	49.8	+0.1	21.9	21.5
75.1	±1.3	75.3	+0.2	21.6	21.3

## Temperature Procedure RUKP2

The probe was calibrated by comparison with platinum resistance thermometers, which are traceable to national standards, and the values taken from the HW4. The calibration was conducted in a temperature chamber in controlled laboratory conditions 23 °C ± 2 °C. The probe under calibration was fully immersed in the chamber. The temperature scale used is ITS-90.

Applied Temperature (°C)	Indicated Temperature (°C)	Instrument Error (°C)	Calibration Uncertainty (°C)**
-0.12	0.02	+0.14	±0.26
40.04	40.02	-0.02	±0.17

\* Not included within the scope of the UKAS accreditation

\*\*The uncertainties quoted apply only to values obtained during the calibration and are not indicative of long-term stability of the instrument under calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. (TSDC39 Issue 1)