



# Manufacturer Certificate of Validation

<b>Certificate No</b>	<b>Date</b>	<b>Page</b>
VC04635EE2EC5A80	23-Oct-19	1/1

imec Messtechnik GmbH certifies that the products mentioned on the following page(s) have been thoroughly tested, validated and met performance accuracy specifications over the stated ranges

## Reference Instrumentation

Brand	Model	Calibration Date	Accuracy/Volatility	Serial Number
Jung	G302	18-Jun-2019	±0.10°C	1734-0001

All reference instruments are calibrated annually by certified laboratories

## Environment Conditions

### Air Temperature

25°C

### Relative Humidity

50 %rh

## Temperature Validation

Validation Points	Tolerance	Result
+60°C	±0.2°C	PASS
+40°C	±0.2°C	PASS
+25°C	±0.2°C	PASS
0°C	±0.2°C	PASS
-5°C	±0.2°C	PASS
-25°C	±0.2°C	PASS

## Product Information

<b>Product Name</b>	<b>Serial Number</b>	<b>Production Date</b>
tempmate.®-S1	04635EE2EC5A80	23-Oct-19

This certificate was automatically generated and is valid without signature.

