

# CERTIFICATE OF CALIBRATION

ISSUED BY AVON-DYNAMIC CALIBRATION

Date of Issue 11 February 2021

Certificate Number K563757



Page 1 of 3

CALIBRATE MEASURE INNOVATE

**For** Signatrol Ltd  
Unit E2, Green Lane Business Park  
Tewkesbury  
Glos.  
GL20 8SJ

**Approved Signatory:**  
M.Hyde

---

<u>Customer Ref Number</u> :	CE1159	<u>Date of Calibration</u> :	11 February 2021
<u>Date of Receipt</u> :	8 February 2021	<u>Item Type</u> :	SDM3055
<u>Item Serial Number</u> :	SDM35FAQ1R1792		
<u>Instrument Manufacturer</u> :	Siglent		
<u>Description</u> :	Digital Multimeter		
<u>Specification Reference</u> :	Manufacturer's Stated Specification (Manual Ref.No.34410-90010)		
<u>Procedure Reference</u> :	CLI050, CLI051, CLI052, CLI053, CL1054 & CLI055.		

## Note

**Calibrations marked ## (Not UKAS Accredited) in this Certificate have been included for Completeness**

**Calibrated By :** M.Hyde

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $\kappa = 2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

---

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory

---

# CERTIFICATE OF CALIBRATION

ISSUED BY AVON-DYNAMIC CALIBRATION

UKAS ACCREDITED CALIBRATION LABORATORY No 0199

Certificate Number  
K563757  
Page 2 of 3

## REPORT

The unit under test was allowed to stabilise for 24 hours in the laboratory environment prior to testing. The unit was allowed to settle for 1 minute before each reading was taken. Results: As received. No adjustment was necessary.

### DC Voltage Ranges (100mV to 1 000V)

<u>Range</u>	<u>Applied Voltage</u>	<u>Instrument Reading</u>	<u>Specification</u>	<u>Uncertainty of Measurement</u>
200 mV	+ 100.000 00 mV	+ 99.999 mV	± 23 µV	± 1.5 µV
	- 100.000 00 mV	- 99.998 mV	± 23 µV	± 1.5 µV
2 V	+ 1.000 000 0 V	+ 0.999 99 V	± 210 µV	± 8 µV
	- 1.000 000 0 V	- 0.999 99 V	± 210 µV	± 8 µV
20 V	+ 19.000 000 V	+ 19.001 4 V	± 3.7 mV	± 60 µV
	+ 10.000 000 V	+ 10.001 1 V	± 2.3 mV	± 60 µV
	+ 5.000 000 V	+ 5.000 5 V	± 1.6 mV	± 40 µV
	- 10.000 000 V	- 9.999 5 V	± 2.3 mV	± 60 µV
200 V	+ 100.000 00 V	+ 99.999 V	± 21 mV	± 0.80 mV
1 000 V	+ 1 000.000 0 V	+ 999.97 V	± 180 mV	± 9.1 mV

### AC Voltage Ranges (100mV to 750V)

<u>Range</u>	<u>Applied Voltage</u>	<u>Instrument Reading</u>	<u>Specification</u>	<u>Uncertainty of Measurement</u>
200 mV	100.000 0 mV @ 1 kHz	99.917 mV	± 300 µV	± 19.6 µV
	100.000 0 mV @ 30 kHz	100.140 mV	± 1.1 mV	± 32.6 µV
2 V	1.000 000 V @ 1 kHz	0.999 76 V	± 3.0 mV	± 121 µV
	1.000 000 V @ 30 kHz	1.001 41 V	± 11 mV	± 253 µV
20 V	10.000 00 V @ 1 kHz	9.984 4 V	± 30 mV	± 0.62 mV
	10.000 00 V @ 30 kHz	10.025 8 V	± 0.11 V	± 1.03 mV
200 V	100.000 0 V @ 1 kHz	99.835 V	± 1.2 V	± 7.2 mV
	100.000 0 V @ 30 kHz	100.249 V	± 1.2 V	± 11.0 mV
750 V	750.000 V @ 1 kHz	748.876 V	± 1.85 V	± 79 mV
	210.000 V @ 30 kHz	210.567 V	± 2.48 V	± 22 mV

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $\kappa = 2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory

# CERTIFICATE OF CALIBRATION

ISSUED BY AVON-DYNAMIC CALIBRATION

UKAS ACCREDITED CALIBRATION LABORATORY No 0199

Certificate Number  
K563757  
Page 3 of 3

## Resistance Ranges (100Ω to 100MΩ)

<u>Range</u>	<u>Applied Resistance</u>	<u>Instrument Reading</u>	<u>Specification</u>	<u>Uncertainty of Measurement</u>
200 Ω	99.997 79 Ω	100.018 Ω	± 40 mΩ	± 1.4 mΩ
2 kΩ	0.999 929 kΩ	1.000 07 kΩ	± 260 mΩ	± 12 mΩ
20 kΩ	9.999 427 kΩ	10.000 7 kΩ	± 2.6 Ω	± 120 mΩ
200 kΩ	99.994 16 kΩ	100.007 kΩ	± 40 Ω	± 1.5 Ω
2 MΩ	0.999 8858 MΩ	0.999 95 MΩ	± 480 Ω	± 25 Ω
10 MΩ	9.999 492 MΩ	10.000 1 MΩ	± 25.3 kΩ	± 480 Ω
100 MΩ	100.008 44 MΩ	100.053 MΩ	± 1.754 MΩ	± 12.5 kΩ

## DC Current Range (10mA, 100mA 1A & 3A)

<u>Range</u>	<u>Applied Voltage</u>	<u>Instrument Reading</u>	<u>Specification</u>	<u>Uncertainty of Measurement</u>
200 μA	+ 100.000 0 μA	+ 99.993 μA	± 0.065 μA	± 11.8 nA
2 mA	+ 1.000 000 mA	+ 0.999 94 mA	± 0.65 μA	± 51 nA
20 mA	+ 10.000 00 mA	+ 9.996 9 mA	± 13.5 μA	± 0.48 μA
200mA	+ 100.000 0 mA	+ 99.988 mA	± 86 μA	± 6.3 μA
2 A	+ 1.000 000 A	+ 0.998 82 A	± 2.1 mA	± 108 μA
10 A	+ 2.000 000 A	+ 1.997 0 A	± 6.0 mA	± 209 μA

## AC Current Range (1A & 3A)

<u>Range</u>	<u>Applied Voltage</u>	<u>Instrument Reading</u>	<u>Specification</u>	<u>Uncertainty of Measurement</u>
20 mA	10.000 00 mA @ 1kHz	10.000 1 mA	± 70 μA	± 0.48 μA
200mA	100.000 0 mA @ 1kHz	99.964 mA	± 1.2 mA	± 6.3 μA
2 A	1.000 000 A @ 1kHz	0.998 94 A	± 9.0 mA	± 344 μA
10 A	2.000 000 A @ 1kHz	1.997 6 A	± 10 mA	± 650 μA

## Frequency

<u>Range</u>	<u>Applied Input</u>	<u>Instrument Reading</u>	<u>Uncertainty of Measurement</u>
100 mV	100.000 00 mV @ 100 Hz	99.998 Hz	± 0.1 Hz
1 V	1.000 000 V @ 100 kHz	99.998 kHz	± 10 Hz

Standards Used  
ADC2613

Laboratory Ambient Temperature = 20°C ± 3°C

Laboratory Humidity = 50% rh ± 20% rh

-End of Report-

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory